

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN072972.D
 Acq On : 21 Jun 2022 10:19
 Operator : JC\MD
 Sample : VN0621WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0621WBL01

Quant Time: Jun 22 06:58:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	338709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	547298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	485520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.609	152	228835	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	199850	43.642	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.280%
35) Dibromofluoromethane	7.951	113	173500	51.289	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.580%
50) Toluene-d8	10.374	98	555497	44.131	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.260%
62) 4-Bromofluorobenzene	12.668	95	238728	52.743	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN072972.D
Acq On : 21 Jun 2022 10:19
Operator : JC\MD
Sample : VN0621WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0621WBL01

Quant Time: Jun 22 06:58:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

