

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN072977.D
 Acq On : 21 Jun 2022 12:40
 Operator : JC\MD
 Sample : N3357-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-2022-06-14

Quant Time: Jun 22 06:59:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	328440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	523830	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	482753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	240535	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	192125	43.267	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.540%
35) Dibromofluoromethane	7.951	113	166243	51.345	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.700%
50) Toluene-d8	10.375	98	538901	44.731	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.460%
62) 4-Bromofluorobenzene	12.669	95	240414	55.495	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN072977.D
Acq On : 21 Jun 2022 12:40
Operator : JC\MD
Sample : N3357-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-2022-06-14

Quant Time: Jun 22 06:59:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

