

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN072982.D
 Acq On : 21 Jun 2022 14:42
 Operator : JC\MD
 Sample : N3379-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-2022-06-16

Quant Time: Jun 22 07:00:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	319799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	512872	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	459237	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.609	152	220841	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	188959	43.704	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.400%
35) Dibromofluoromethane	7.951	113	164499	51.892	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.780%
50) Toluene-d8	10.374	98	529438	44.884	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.760%
62) 4-Bromofluorobenzene	12.668	95	228120	53.782	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.560%
Target Compounds						
17) Carbon Disulfide	4.463	76	79785	10.170	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN072982.D
Acq On : 21 Jun 2022 14:42
Operator : JC\MD
Sample : N3379-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-2022-06-16

Quant Time: Jun 22 07:00:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

