

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN072985.D
 Acq On : 21 Jun 2022 15:55
 Operator : JC\MD
 Sample : N3379-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-04

Quant Time: Jun 22 07:01:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	302234	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	492499	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	448761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	224429	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	182050	44.553	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.100%
35) Dibromofluoromethane	7.951	113	154788	50.849	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.700%
50) Toluene-d8	10.375	98	509201	44.954	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.900%
62) 4-Bromofluorobenzene	12.663	95	219849	53.976	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.960%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.410	62	19185	7.434	ug/l #	84
19) Methyl tert-butyl Ether	5.545	73	40392	3.717	ug/l	92
27) cis-1,2-Dichloroethene	7.251	96	26581	7.068	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN072985.D
Acq On : 21 Jun 2022 15:55
Operator : JC\MD
Sample : N3379-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-04

Quant Time: Jun 22 07:01:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

