

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN072992.D
 Acq On : 21 Jun 2022 18:45
 Operator : JC\MD
 Sample : N3381-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3

Quant Time: Jun 22 07:02:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	302409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	483852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	428209	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	208453	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	180473	44.142	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.280%
35) Dibromofluoromethane	7.951	113	153741	51.407	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.820%
50) Toluene-d8	10.375	98	498522	44.798	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.600%
62) 4-Bromofluorobenzene	12.669	95	213292	53.302	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.600%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	95612	42.543	ug/l	93
17) Carbon Disulfide	4.463	76	3735	0.503	ug/l #	87
25) 2-Butanone	7.275	43	9967	2.795	ug/l	94
31) Cyclohexane	8.016	56	10613	1.919	ug/l #	51
39) Methylcyclohexane	9.386	83	1059	0.217	ug/l	94
40) Benzene	8.392	78	125120	9.838	ug/l	100
52) Toluene	10.445	92	32159	4.119	ug/l	99
67) Ethyl Benzene	11.780	91	124470	8.455	ug/l	98
68) m/p-Xylenes	11.892	106	79981	13.852	ug/l	98
69) o-Xylene	12.216	106	56664	9.821	ug/l	94
73) Isopropylbenzene	12.516	105	4469	0.277	ug/l	99
78) n-propylbenzene	12.857	91	6576	0.371	ug/l	98
80) 1,3,5-Trimethylbenzene	12.998	105	6515	0.492	ug/l	99
84) 1,2,4-Trimethylbenzene	13.304	105	37979	2.879	ug/l	100
95) Naphthalene	15.427	128	50734	3.151	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN072992.D
Acq On : 21 Jun 2022 18:45
Operator : JC\MD
Sample : N3381-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3

Quant Time: Jun 22 07:02:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

