

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN072993.D
 Acq On : 21 Jun 2022 19:10
 Operator : JC\MD
 Sample : N3381-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-6

Quant Time: Jun 22 07:02:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	317782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	508972	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	471201	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	228137	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	191425	44.555	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.120%
35) Dibromofluoromethane	7.951	113	163081	51.839	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.680%
50) Toluene-d8	10.375	98	534588	45.668	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.340%
62) 4-Bromofluorobenzene	12.669	95	231015	54.882	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.760%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	82829	35.072	ug/l	97
25) 2-Butanone	7.281	43	6841	1.825	ug/l	91
67) Ethyl Benzene	11.780	91	3781	0.233	ug/l #	75
68) m/p-Xylenes	11.892	106	8677	1.366	ug/l	95
69) o-Xylene	12.222	106	1654	0.261	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN072993.D
Acq On : 21 Jun 2022 19:10
Operator : JC\MD
Sample : N3381-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-6

Quant Time: Jun 22 07:02:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

