

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN073004.D
 Acq On : 22 Jun 2022 05:17
 Operator : JC\MD
 Sample : N3381-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-27

Quant Time: Jun 22 07:04:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	288931	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	462947	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.681	117	419404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	207190	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	173019	44.293	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.580%
35) Dibromofluoromethane	7.951	113	149196	52.140	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.280%
50) Toluene-d8	10.375	98	474924	44.605	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.200%
62) 4-Bromofluorobenzene	12.669	95	208151	54.366	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.740%
Target Compounds						
16) Acetone	4.228	43	45405	21.146	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN073004.D
Acq On : 22 Jun 2022 05:17
Operator : JC\MD
Sample : N3381-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-27

Quant Time: Jun 22 07:04:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

