

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN073005.D
 Acq On : 22 Jun 2022 05:41
 Operator : JC\MD
 Sample : N3381-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-34

Quant Time: Jun 22 07:04:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	309311	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	478111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	457461	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	243262	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.381	65	161623	38.649	ug/l	0.01
Spiked Amount 50.000	Range 74 - 125		Recovery =	77.300%		
35) Dibromofluoromethane	7.951	113	154669	52.339	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	104.680%		
50) Toluene-d8	10.375	98	509403	46.325	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.660%		
62) 4-Bromofluorobenzene	12.669	95	238771	60.386	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	120.780%		
Target Compounds					Qvalue	
16) Acetone	4.228	43	84201	36.630	ug/l	94
22) Diisopropyl ether	6.416	45	61999	5.256	ug/l #	88
31) Cyclohexane	8.022	56	122361	21.631	ug/l	93
39) Methylcyclohexane	9.392	83	10687	2.217	ug/l	90
40) Benzene	8.375	78	33818478m	2691.088	ug/l	
52) Toluene	10.439	92	1694789	219.664	ug/l	99
67) Ethyl Benzene	11.780	91	14034907	892.407	ug/l	98
68) m/p-Xylenes	11.892	106	12958504	2100.793	ug/l	84
69) o-Xylene	12.216	106	2038725	330.763	ug/l	95
73) Isopropylbenzene	12.516	105	218938	11.634	ug/l	99
78) n-propylbenzene	12.857	91	377290	18.226	ug/l	97
80) 1,3,5-Trimethylbenzene	12.998	105	552629	35.749	ug/l	98
84) 1,2,4-Trimethylbenzene	13.304	105	2517549	163.509	ug/l	97
95) Naphthalene	15.427	128	1316789	70.070	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN073005.D
Acq On : 22 Jun 2022 05:41
Operator : JC\MD
Sample : N3381-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-34

Quant Time: Jun 22 07:04:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

