

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN073007.D
 Acq On : 22 Jun 2022 06:30
 Operator : JC\MD
 Sample : N3381-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PZ-10

Quant Time: Jun 22 07:05:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	305471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	490535	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	444177	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	217258	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	181705	43.998	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.000%
35) Dibromofluoromethane	7.951	113	157081	51.809	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.620%
50) Toluene-d8	10.374	98	507461	44.980	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.960%
62) 4-Bromofluorobenzene	12.668	95	219079	54.002	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.000%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	40011	17.625	ug/l	99
40) Benzene	8.392	78	237369	18.410	ug/l	96
52) Toluene	10.439	92	2951	0.373	ug/l	92
67) Ethyl Benzene	11.780	91	14329	0.938	ug/l	96
68) m/p-Xylenes	11.886	106	18293	3.054	ug/l	99
95) Naphthalene	15.433	128	15225	0.907	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN073007.D
Acq On : 22 Jun 2022 06:30
Operator : JC\MD
Sample : N3381-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PZ-10

Quant Time: Jun 22 07:05:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

