

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN073008.D
 Acq On : 22 Jun 2022 06:54
 Operator : JC\MD
 Sample : N3381-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2022-06-16

Quant Time: Jun 22 07:22:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	294534	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	476926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	430193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	198606	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	176842	44.410	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.820%
35) Dibromofluoromethane	7.951	113	153815	52.179	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.360%
50) Toluene-d8	10.375	98	494532	45.085	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.160%
62) 4-Bromofluorobenzene	12.668	95	208386	52.832	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.660%
Target Compounds						
					Qvalue	
40) Benzene	8.386	78	10372	0.827	ug/l #	87
67) Ethyl Benzene	11.780	91	5888	0.398	ug/l	94
68) m/p-Xylenes	11.886	106	6788	1.170	ug/l	92
95) Naphthalene	15.427	128	6528	0.425	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN073008.D
Acq On : 22 Jun 2022 06:54
Operator : JC\MD
Sample : N3381-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2022-06-16

Quant Time: Jun 22 07:22:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

