

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN073015.D
 Acq On : 22 Jun 2022 09:44
 Operator : JC\MD
 Sample : N3334-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-11

Quant Time: Jun 23 16:57:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	306837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	500502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	448486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	215714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	186835	45.038	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.080%
35) Dibromofluoromethane	7.951	113	161221	52.115	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.240%
50) Toluene-d8	10.374	98	520983	45.259	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.520%
62) 4-Bromofluorobenzene	12.668	95	219569	53.045	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.100%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	65014	28.511	ug/l	95
20) Methylene Chloride	5.022	84	37105	10.299	ug/l	92
25) 2-Butanone	7.269	43	27081	7.484	ug/l #	83
43) Isopropyl Acetate	8.492	43	29288	3.060	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN073015.D
Acq On : 22 Jun 2022 09:44
Operator : JC\MD
Sample : N3334-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-11

Quant Time: Jun 23 16:57:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

