

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN073016.D
 Acq On : 22 Jun 2022 10:08
 Operator : JC\MD
 Sample : N3299-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COM-4

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/23/2022
 Supervised By :Semsettin Yesilyurt 06/23/2022

Quant Time: Jun 23 16:57:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	26476	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.904	114	81819	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	45705	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	12142	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	163023	455.437	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	910.880%#
35) Dibromofluoromethane	7.945	113	663m	1.311	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	2.620%#
50) Toluene-d8	10.375	98	69910	37.151	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	74.300%#
62) 4-Bromofluorobenzene	12.669	95	18859	27.871	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	55.740%#
Target Compounds						
16) Acetone	4.228	43	96346	489.658	ug/l	Qvalue 94
20) Methylene Chloride	5.016	84	16267	53.316	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN073016.D
Acq On : 22 Jun 2022 10:08
Operator : JC\MD
Sample : N3299-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COM-4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/23/2022
Supervised By :Semsettin Yesilyurt 06/23/2022

Quant Time: Jun 23 16:57:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

