

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN073017.D
 Acq On : 22 Jun 2022 10:33
 Operator : JC\MD
 Sample : N3304-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-10

Quant Time: Jun 23 16:57:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	26426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	75886	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	48730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	11759	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	127360	356.478	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 712.960%#		
35) Dibromofluoromethane	7.951	113	69772	148.753	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 297.500%#		
50) Toluene-d8	10.380	98	52957	30.342	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 60.680%#		
62) 4-Bromofluorobenzene	12.668	95	16262	25.912	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 51.820%#		
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	71830	365.751	ug/l	92
18) Methyl Acetate	4.781	43	247681	473.599	ug/l #	90
20) Methylene Chloride	5.022	84	20879	68.631	ug/l	93
25) 2-Butanone	7.263	43	24058	77.198	ug/l	89
43) Isopropyl Acetate	8.498	43	20382	14.046	ug/l #	87
59) 2-Hexanone	11.027	43	158579	216.668	ug/l #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN073017.D
Acq On : 22 Jun 2022 10:33
Operator : JC\MD
Sample : N3304-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-10

Quant Time: Jun 23 16:57:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

