

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN073019.D
 Acq On : 22 Jun 2022 11:21
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 23 16:58:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.476	0.527	-10.7	81	0.00
3 P	Chloromethane	0.535	0.507	5.2	78	0.00
4 C	Vinyl Chloride	0.427	0.471	-10.3#	82	0.00
5 T	Bromomethane	0.168	0.168	0.0	70	0.00
6 T	Chloroethane	0.267	0.302	-13.1	87	0.00
7 T	Trichlorofluoromethane	0.695	0.881	-26.8#	104	0.01
8 T	Diethyl Ether	0.359	0.389	-8.4	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.481	0.518	-7.7	83	0.00
10 T	Methyl Iodide	0.551	0.406	26.3#	56	0.00
11 T	Tert butyl alcohol	0.175	0.176	-0.6	75	-0.01
12 CM	1,1-Dichloroethene	0.479	0.499	-4.2#	79	0.00
13 T	Acrolein	0.069	0.067	2.9	74	0.00
14 T	Allyl chloride	1.023	1.007	1.6	76	0.00
15 T	Acrylonitrile	0.403	0.464	-15.1	87	0.00
16 T	Acetone	0.372	0.413	-11.0	87	0.00
17 T	Carbon Disulfide	1.227	1.077	12.2	68	0.00
18 T	Methyl Acetate	0.990	1.101	-11.2	89	0.00
19 T	Methyl tert-butyl Ether	1.798	2.075	-15.4	86	0.00
20 T	Methylene Chloride	0.628	0.631	-0.5	82	0.00
21 T	trans-1,2-Dichloroethene	0.521	0.534	-2.5	79	0.00
22 T	Diisopropyl ether	1.907	2.270	-19.0	88	0.00
23 T	Vinyl Acetate	1.760	2.084	-18.4	87	0.00
24 P	1,1-Dichloroethane	1.000	1.142	-14.2	86	0.00
25 T	2-Butanone	0.590	0.659	-11.7	85	0.00
26 T	2,2-Dichloropropane	0.898	0.441	50.9#	39#	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.695	-11.7	86	0.00
28 T	Bromochloromethane	0.413	0.479	-16.0	85	0.00
29 T	Tetrahydrofuran	0.389	0.443	-13.9	87	0.00
30 C	Chloroform	1.012	1.155	-14.1#	86	0.00
31 T	Cyclohexane	0.914	0.908	0.7	81	0.00
32 T	1,1,1-Trichloroethane	0.888	1.003	-13.0	84	0.00
33 S	1,2-Dichloroethane-d4	0.676	0.793	-17.3	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.309	0.340	-10.0	88	0.00
36 T	1,1-Dichloropropene	0.418	0.426	-1.9	82	0.00
37 T	Ethyl Acetate	0.669	0.723	-8.1	88	0.00
38 T	Carbon Tetrachloride	0.440	0.451	-2.5	82	0.00
39 T	Methylcyclohexane	0.504	0.476	5.6	75	0.00
40 TM	Benzene	1.314	1.378	-4.9	84	0.00
41 T	Methacrylonitrile	0.317	0.358	-12.9	93	0.00
42 TM	1,2-Dichloroethane	0.467	0.510	-9.2	86	0.00
43 T	Isopropyl Acetate	0.956	1.069	-11.8	88	0.00
44 TM	Trichloroethene	0.333	0.330	0.9	81	0.00
45 C	1,2-Dichloropropane	0.336	0.371	-10.4#	89	0.00
46 T	Dibromomethane	0.227	0.247	-8.8	86	0.00
47 T	Bromodichloromethane	0.451	0.504	-11.8	88	0.00
48 T	Methyl methacrylate	0.437	0.480	-9.8	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN073019.D
 Acq On : 22 Jun 2022 11:21
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 23 16:58:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	76	0.00
50 S	Toluene-d8	1.150	1.277	-11.0	88	0.00
51 T	4-Methyl-2-Pentanone	0.623	0.707	-13.5	89	0.00
52 CM	Toluene	0.807	0.869	-7.7#	85	0.00
53 T	t-1,3-Dichloropropene	0.485	0.487	-0.4	75	0.00
54 T	cis-1,3-Dichloropropene	0.530	0.524	1.1	77	0.00
55 T	1,1,2-Trichloroethane	0.354	0.368	-4.0	87	0.00
56 T	Ethyl methacrylate	0.568	0.638	-12.3	86	0.00
57 T	1,3-Dichloropropane	0.588	0.630	-7.1	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.251	0.284	-13.1	83	0.00
59 T	2-Hexanone	0.482	0.540	-12.0	87	0.00
60 T	Dibromochloromethane	0.335	0.384	-14.6	85	0.00
61 T	1,2-Dibromoethane	0.349	0.385	-10.3	87	0.00
62 S	4-Bromofluorobenzene	0.414	0.464	-12.1	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.304	0.293	3.6	80	0.00
65 PM	Chlorobenzene	0.976	1.009	-3.4	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.389	-4.6	84	0.00
67 C	Ethyl Benzene	1.719	1.856	-8.0#	85	0.00
68 T	m/p-Xylenes	0.674	0.712	-5.6	85	0.00
69 T	o-Xylene	0.674	0.717	-6.4	86	0.00
70 T	Styrene	1.082	1.187	-9.7	86	0.00
71 P	Bromoform	0.280	0.316	-12.9	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	3.868	4.088	-5.7	86	0.00
74 T	N-amyl acetate	2.007	2.181	-8.7	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.393	1.482	-6.4	89	0.00
76 T	1,2,3-Trichloropropane	1.212	1.349	-11.3	102	0.00
77 T	Bromobenzene	0.927	0.951	-2.6	83	0.00
78 T	n-propylbenzene	4.255	4.664	-9.6	86	0.00
79 T	2-Chlorotoluene	2.755	2.843	-3.2	85	0.00
80 T	1,3,5-Trimethylbenzene	3.177	3.373	-6.2	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.449	0.379	15.6	65	0.00
82 T	4-Chlorotoluene	2.607	2.769	-6.2	85	0.00
83 T	tert-Butylbenzene	2.910	3.129	-7.5	86	0.00
84 T	1,2,4-Trimethylbenzene	3.165	3.395	-7.3	86	0.00
85 T	sec-Butylbenzene	3.882	4.208	-8.4	86	0.00
86 T	p-Isopropyltoluene	3.143	3.396	-8.0	84	0.00
87 T	1,3-Dichlorobenzene	1.654	1.708	-3.3	85	0.00
88 T	1,4-Dichlorobenzene	1.674	1.727	-3.2	85	0.00
89 T	n-Butylbenzene	2.566	2.783	-8.5	83	0.00
90 T	Hexachloroethane	0.592	0.672	-13.5	89	0.00
91 T	1,2-Dichlorobenzene	1.725	1.781	-3.2	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.376	0.362	3.7	83	0.00
93 T	1,2,4-Trichlorobenzene	0.955	0.956	-0.1	81	0.00
94 T	Hexachlorobutadiene	0.416	0.384	7.7	75	0.00
95 T	Naphthalene	3.863	3.855	0.2	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN073019.D
Acq On : 22 Jun 2022 11:21
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 23 16:58:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.020	1.007	1.3	84	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6