

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN072998.D
 Acq On : 22 Jun 2022 02:51
 Operator : JC\MD
 Sample : VN0621WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0621WBL02

Quant Time: Jun 22 07:03:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	339216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	539411	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	483798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	230491	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	200633	43.748	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.500%
35) Dibromofluoromethane	7.951	113	172504	51.740	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.480%
50) Toluene-d8	10.374	98	558823	45.044	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.080%
62) 4-Bromofluorobenzene	12.668	95	238690	53.505	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.020%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN072998.D
Acq On : 22 Jun 2022 02:51
Operator : JC\MD
Sample : VN0621WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0621WBL02

Quant Time: Jun 22 07:03:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.016 min Scan# 1
Delta R.T. -0.000 min
Lab File: VN072998.D
Acq: 22 Jun 2022 02:51




Instrument :
MSVOA_N
ClientSampleId :
VN0621WBL02


Tgt Ion:168 Resp: 339216
Ion Ratio Lower Upper
168 100
99 59.6 42.3 63.5



#33
1,2-Dichloroethane-d4
Concen: 43.748 ug/l
RT: 8.369 min Scan# 1102
Delta R.T. -0.000 min
Lab File: VN072998.D
Acq: 22 Jun 2022 02:51

Tgt Ion: 65 Resp: 200633
Ion Ratio Lower Upper
65 100
67 51.6 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.904 min Scan# 1192

Delta R.T. 0.006 min

Lab File: VN072998.D

Acq: 22 Jun 2022 02:51

Instrument :

MSVOA_N

ClientSampleId :

VN0621WBL02





Tgt Ion:114 Resp: 539411

Ion Ratio Lower Upper

114 100

63 21.1 0.0 40.0

88 14.2 0.0 32.6



#35

Dibromofluoromethane

Concen: 51.740 ug/l

RT: 7.951 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VN072998.D

Acq: 22 Jun 2022 02:51




Tgt Ion:113 Resp: 172504


Ion Ratio Lower Upper

113 100

111 104.4 81.4 122.0

192 18.8 17.0 25.6









#50
Toluene-d8
Concen: 45.044 ug/l
RT: 10.374 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN072998.D
Acq: 22 Jun 2022 02:51




Instrument :
MSVOA_N
ClientSampleId :
VN0621WBL02


Tgt Ion: 98 Resp: 558823
Ion Ratio Lower Upper
98 100
100 65.6 52.6 78.8



#62
4-Bromofluorobenzene
Concen: 53.505 ug/l
RT: 12.668 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN072998.D
Acq: 22 Jun 2022 02:51

Tgt Ion: 95 Resp: 238690
Ion Ratio Lower Upper
95 100
174 75.8 0.0 167.6
176 73.2 0.0 163.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.680 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN072998.D
Acq: 22 Jun 2022 02:51

Instrument :
MSVOA_N
ClientSampleId :
VN0621WBL02

Tgt Ion:117 Resp: 483798
Ion Ratio Lower Upper
117 100
82 55.3 44.2 66.2
119 31.5 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.610 min Scan# 1993
Delta R.T. -0.000 min
Lab File: VN072998.D
Acq: 22 Jun 2022 02:51

Tgt Ion:152 Resp: 230491
Ion Ratio Lower Upper
152 100
115 59.5 28.9 86.7
150 159.3 0.0 356.2

