



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN072969.D
 Acq On : 21 Jun 2022 08:30
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Title : SW846 8260
 Last Update : Sat Jun 18 06:22:45 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1824

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	12268	PASS
75	95	30	60	51.0	31195	PASS
95	95	100	100	100.0	61112	PASS
96	95	5	9	6.6	4056	PASS
173	174	0.00	2	0.9	411	PASS
174	95	50	100	77.4	47331	PASS
175	174	5	9	7.7	3632	PASS
176	174	95	101	97.1	45947	PASS
177	176	5	9	7.1	3252	PASS