

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078316.D
 Acq On : 21 Jun 2023 12:35
 Operator : JC\MD
 Sample : VN0621MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0621MBL01

Quant Time: Jun 22 04:46:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	433875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	786767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	681072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	194718	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	306360	52.959	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.920%
35) Dibromofluoromethane	8.177	113	278847	56.054	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.100%
50) Toluene-d8	10.571	98	906425	47.265	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.520%
62) 4-Bromofluorobenzene	12.853	95	276189	46.326	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078316.D
Acq On : 21 Jun 2023 12:35
Operator : JC\MD
Sample : VN0621MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0621MBL01

Quant Time: Jun 22 04:46:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

