

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078317.D
 Acq On : 21 Jun 2023 12:59
 Operator : JC\MD
 Sample : VN0621WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0621WBL01

Quant Time: Jun 22 04:46:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	528182	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	987912	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	884458	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	293449	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	377482	53.602	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.200%
35) Dibromofluoromethane	8.177	113	351115	56.210	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.420%
50) Toluene-d8	10.571	98	1106549	45.952	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.900%
62) 4-Bromofluorobenzene	12.853	95	383216	51.190	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078317.D
Acq On : 21 Jun 2023 12:59
Operator : JC\MD
Sample : VN0621WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0621WBL01

Quant Time: Jun 22 04:46:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

