

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078324.D
 Acq On : 21 Jun 2023 15:58
 Operator : JC\MD
 Sample : PB153609TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153609TB

Quant Time: Jun 22 04:50:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	455552	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	844453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	803236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	262200	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	324884	53.489	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.980%			
35) Dibromofluoromethane	8.177	113	294956	55.242	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.480%			
50) Toluene-d8	10.571	98	1026852	49.886	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.780%			
62) 4-Bromofluorobenzene	12.853	95	336176	52.536	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 105.080%			
Target Compounds					Qvalue	
16) Acetone	4.453	43	24168	13.824	ug/l	92
20) Methylene Chloride	5.289	84	21287	3.685	ug/l #	76
37) Ethyl Acetate	7.577	43	61278	10.242	ug/l	96
43) Isopropyl Acetate	8.694	43	284482	24.863	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078324.D
Acq On : 21 Jun 2023 15:58
Operator : JC\MD
Sample : PB153609TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153609TB

Quant Time: Jun 22 04:50:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

