

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078325.D
 Acq On : 21 Jun 2023 16:22
 Operator : JC\MD
 Sample : 03336-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-6

Quant Time: Jun 22 04:51:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	473649	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	863595	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	811200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	254480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	331052	52.422	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.840%		
35) Dibromofluoromethane	8.177	113	301183	55.158	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 110.320%		
50) Toluene-d8	10.571	98	1039686	49.390	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 98.780%		
62) 4-Bromofluorobenzene	12.853	95	333787	51.006	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 102.020%		
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	25518	14.038	ug/l #	84
20) Methylene Chloride	5.294	84	24193	4.028	ug/l	94
37) Ethyl Acetate	7.577	43	57320	9.368	ug/l #	95
43) Isopropyl Acetate	8.694	43	276837	23.658	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078325.D
Acq On : 21 Jun 2023 16:22
Operator : JC\MD
Sample : 03336-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-6

Quant Time: Jun 22 04:51:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

