

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078327.D
 Acq On : 21 Jun 2023 17:10
 Operator : JC\MD
 Sample : 03336-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-4

Quant Time: Jun 22 04:52:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	459241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	866972	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	814077	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	260720	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	327095	53.420	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.840%	
35) Dibromofluoromethane	8.177	113	303637	55.391	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.780%	
50) Toluene-d8	10.576	98	1040418	49.233	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.460%	
62) 4-Bromofluorobenzene	12.853	95	337943	51.440	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.880%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	38811	22.021	ug/l	95
20) Methylene Chloride	5.294	84	23571	4.047	ug/l	92
37) Ethyl Acetate	7.571	43	57650	9.385	ug/l #	94
43) Isopropyl Acetate	8.700	43	317043	26.989	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078327.D
Acq On : 21 Jun 2023 17:10
Operator : JC\MD
Sample : 03336-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-4

Quant Time: Jun 22 04:52:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

