

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078329.D
 Acq On : 21 Jun 2023 17:59
 Operator : JC\MD
 Sample : 03336-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2

Quant Time: Jun 22 04:53:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	440018	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	843326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	796882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	262283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	312916	53.337	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.680%		
35) Dibromofluoromethane	8.171	113	290370	54.456	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.920%		
50) Toluene-d8	10.571	98	1001740	48.731	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 97.460%		
62) 4-Bromofluorobenzene	12.853	95	336906	52.720	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 105.440%		
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	25727	15.235	ug/l	99
20) Methylene Chloride	5.289	84	24433	4.378	ug/l #	87
37) Ethyl Acetate	7.571	43	60253	10.084	ug/l	95
43) Isopropyl Acetate	8.700	43	419394	36.702	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078329.D
Acq On : 21 Jun 2023 17:59
Operator : JC\MD
Sample : 03336-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2

Quant Time: Jun 22 04:53:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

