

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078330.D
 Acq On : 21 Jun 2023 18:23
 Operator : JC\MD
 Sample : 03336-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Jun 22 04:53:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	448594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	872487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	810550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	268052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	329174	55.035	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.080%	
35) Dibromofluoromethane	8.171	113	299510	54.292	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	108.580%	
50) Toluene-d8	10.570	98	1043984	49.089	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.180%	
62) 4-Bromofluorobenzene	12.853	95	343040	51.886	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	103.780%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	27922	16.219	ug/l #	84
20) Methylene Chloride	5.288	84	24261	4.264	ug/l	92
37) Ethyl Acetate	7.577	43	60679	9.816	ug/l #	96
43) Isopropyl Acetate	8.700	43	439632	37.188	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078330.D
Acq On : 21 Jun 2023 18:23
Operator : JC\MD
Sample : 03336-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Jun 22 04:53:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

