

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078333.D
 Acq On : 21 Jun 2023 19:35
 Operator : JC\MD
 Sample : 03337-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-9

Quant Time: Jun 22 04:54:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	453342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	853319	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	810462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	264272	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	319267	52.820	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.640%		
35) Dibromofluoromethane	8.177	113	296053	54.871	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 109.740%		
50) Toluene-d8	10.577	98	1003909	48.265	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 96.540%		
62) 4-Bromofluorobenzene	12.853	95	338078	52.284	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.560%		
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	26126	15.017	ug/l	99
20) Methylene Chloride	5.295	84	23259	4.046	ug/l #	86
37) Ethyl Acetate	7.577	43	60977	10.086	ug/l	97
43) Isopropyl Acetate	8.694	43	433189	37.466	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078333.D
Acq On : 21 Jun 2023 19:35
Operator : JC\MD
Sample : 03337-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-9

Quant Time: Jun 22 04:54:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

