

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078334.D
 Acq On : 21 Jun 2023 19:59
 Operator : JC\MD
 Sample : 03337-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-8

Quant Time: Jun 22 04:55:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	451060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	856062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	813056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	263408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	320440	53.282	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.560%		
35) Dibromofluoromethane	8.177	113	299311	55.297	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 110.600%		
50) Toluene-d8	10.571	98	1033252	49.517	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.040%		
62) 4-Bromofluorobenzene	12.853	95	338313	52.153	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.300%		
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	25100	14.500	ug/l #	75
20) Methylene Chloride	5.289	84	24272	4.243	ug/l	89
37) Ethyl Acetate	7.571	43	58062	9.573	ug/l	95
43) Isopropyl Acetate	8.700	43	544079	46.906	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078334.D
Acq On : 21 Jun 2023 19:59
Operator : JC\MD
Sample : 03337-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-8

Quant Time: Jun 22 04:55:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

