

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
 Data File : VN078340.D
 Acq On : 21 Jun 2023 22:23
 Operator : JC\MD
 Sample : 03357-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-SCAN-03-G

Quant Time: Jun 22 04:57:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	454759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	876859	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	804330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	259198	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	333007	54.921	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.840%	
35) Dibromofluoromethane	8.171	113	303711	54.779	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.560%	
50) Toluene-d8	10.571	98	1047510	49.009	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.020%	
62) 4-Bromofluorobenzene	12.853	95	333095	50.130	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.260%	
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	172206	98.671	ug/l	93
18) Methyl Acetate	5.047	43	10194	1.269	ug/l #	83
20) Methylene Chloride	5.283	84	14123	2.449	ug/l #	90
25) 2-Butanone	7.500	43	43178	16.088	ug/l	97
37) Ethyl Acetate	7.577	43	15340	2.469	ug/l #	91
43) Isopropyl Acetate	8.694	43	187657	15.794	ug/l #	75
51) 4-Methyl-2-Pentanone	10.453	43	289334	46.245	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078340.D
Acq On : 21 Jun 2023 22:23
Operator : JC\MD
Sample : 03357-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-SCAN-03-G

Quant Time: Jun 22 04:57:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

