


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062123\
Data File : VN078314.D
Acq On : 21 Jun 2023 11:23
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Title : SW846 8260
Last Update : Wed Jun 07 05:05:00 2023



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1843

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	14246	PASS
75	95	30	60	55.1	42032	PASS
95	95	100	100	100.0	76251	PASS
96	95	5	9	6.9	5241	PASS
173	174	0.00	2	1.2	661	PASS
174	95	50	100	73.5	56013	PASS
175	174	5	9	6.8	3819	PASS
176	174	95	101	99.6	55781	PASS
177	176	5	9	6.2	3434	PASS

Average of 12.847 to 12.859 min.: VN078314.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.90	943	50.00	14246	61.10	3141	71.80	491
37.00	3409	51.00	4774	62.05	3743	73.00	420
38.00	3807	51.95	224	62.95	4017	74.00	1402
38.95	1656	54.70	76	63.95	396	75.00	4203
40.00	596	55.00	179	64.20	113	76.05	3943
43.95	341	56.00	1409	64.90	632	76.85	347
44.95	836	57.00	2683	66.10	96	77.00	157
46.05	183	58.20	103	66.85	237	77.20	144
47.10	574	59.85	508	68.00	8172	77.80	57
47.90	479	60.10	716	69.00	9322	78.05	315
48.95	2954	60.90	1314	70.00	773	78.90	3478

Average of 12.847 to 12.859 min.: VN078314.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
79.85	938	94.05	10503	117.80	561	145.20	64
80.85	3754	95.00	76251	119.00	468	145.80	156
81.85	834	96.00	5241	127.85	302	147.05	130
82.70	78	96.95	154	128.80	131	147.95	242
86.95	3458	103.85	377	129.85	406	152.90	51
87.95	2852	104.80	135	131.00	59	154.75	152
88.70	59	105.85	375	134.80	122	156.00	53
90.50	81	106.70	52	137.00	61	159.10	51
90.90	316	110.80	60	140.90	818	170.75	225
91.95	2665	115.75	451	142.95	1378	171.10	114
92.95	3875	116.80	731	143.90	52	171.95	1136

Average of 12.847 to 12.859 min.: VN078314.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
172.90	661						
173.95	56013						
174.95	3819						
176.00	55781						
176.90	3434						
177.80	60						
281.10	52						