

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062124\
 Data File : VN082681.D
 Acq On : 21 Jun 2024 16:32
 Operator : JC\MD
 Sample : P3004-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-062024

Quant Time: Jun 24 02:14:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	143328	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	262228	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	252386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101282	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	117377	51.068	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.140%
35) Dibromofluoromethane	8.171	113	87166	48.999	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	10.570	98	309504	46.007	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.020%
62) 4-Bromofluorobenzene	12.853	95	95866	41.334	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	82.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062124\
Data File : VN082681.D
Acq On : 21 Jun 2024 16:32
Operator : JC\MD
Sample : P3004-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-062024

Quant Time: Jun 24 02:14:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

