

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062124\
 Data File : VN082682.D
 Acq On : 21 Jun 2024 16:57
 Operator : JC\MD
 Sample : P3004-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWBR-05-145-165-062024

Quant Time: Jun 24 02:14:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	134700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	247627	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	241432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97468	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	116759	54.053	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.100%	
35) Dibromofluoromethane	8.171	113	85739	51.039	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.080%	
50) Toluene-d8	10.570	98	303182	47.725	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.440%	
62) 4-Bromofluorobenzene	12.853	95	97239	44.398	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 88.800%	
Target Compounds						
52) Toluene	10.629	92	15092	3.209	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062124\
Data File : VN082682.D
Acq On : 21 Jun 2024 16:57
Operator : JC\MD
Sample : P3004-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWBR-05-145-165-062024

Quant Time: Jun 24 02:14:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

