

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062124\
 Data File : VN082687.D
 Acq On : 21 Jun 2024 19:00
 Operator : JC\MD
 Sample : P2976-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-K-COMP

Quant Time: Jun 24 02:16:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	130607	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	258425	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	269685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104091	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	117857	56.271	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.540%
35) Dibromofluoromethane	8.171	113	86202	49.170	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.340%
50) Toluene-d8	10.571	98	329983	49.773	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.540%
62) 4-Bromofluorobenzene	12.853	95	103534	45.297	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	90.600%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	32824	34.460	ug/l	97
20) Methylene Chloride	5.271	84	8053	4.809	ug/l #	91
43) Isopropyl Acetate	8.694	43	264684	42.425	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062124\
Data File : VN082687.D
Acq On : 21 Jun 2024 19:00
Operator : JC\MD
Sample : P2976-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-K-COMP

Quant Time: Jun 24 02:16:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

