

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062124\
 Data File : VN082688.D
 Acq On : 21 Jun 2024 19:25
 Operator : JC\MD
 Sample : P2998-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WCS-TPL

Quant Time: Jun 24 02:17:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	132081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	260720	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	274185	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104210	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	117430	55.441	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.880%			
35) Dibromofluoromethane	8.171	113	86012	48.630	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.260%			
50) Toluene-d8	10.571	98	326771	48.855	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.720%			
62) 4-Bromofluorobenzene	12.853	95	101810	44.150	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 88.300%			
Target Compounds					Qvalue	
16) Acetone	4.442	43	31986	33.205	ug/l	99
20) Methylene Chloride	5.277	84	7482	4.418	ug/l #	91
43) Isopropyl Acetate	8.694	43	254842	40.488	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062124\
Data File : VN082688.D
Acq On : 21 Jun 2024 19:25
Operator : JC\MD
Sample : P2998-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WCS-TPL

Quant Time: Jun 24 02:17:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

