

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062218\
 Data File : VN049504.D
 Acq On : 22 Jun 2018 20:41
 Operator : MD\SY
 Sample : J3646-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 22 22:17:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1131040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1810808	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1489695	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	411944	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	780635	52.04	ug/l	0.00
Spiked Amount			Recovery	=	104.08%	
35) Dibromofluoromethane	7.59	113	689014	47.03	ug/l	0.00
Spiked Amount			Recovery	=	94.06%	
50) Toluene-d8	10.09	98	2470967	44.30	ug/l	0.00
Spiked Amount			Recovery	=	88.60%	
62) 4-Bromofluorobenzene	12.41	95	616537	33.43	ug/l	0.00
Spiked Amount			Recovery	=	66.86%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.41	74	7129	0.92	ug/l	99
13) Acrolein	3.61	56	257	0.28	ug/l #	14
14) Allyl chloride	4.55	41	31074	1.52	ug/l #	20
16) Acetone	3.82	43	70360	19.32	ug/l #	88
18) Methyl Acetate	4.33	43	83352	7.51	ug/l	97
20) Methylene Chloride	4.55	84	880368	57.87	ug/l	91
25) 2-Butanone	6.83	43	1175	0.24	ug/l #	23
30) Chloroform	7.37	83	13766	0.53	ug/l	99
31) Cyclohexane	7.66	56	21331	0.79	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	81180	3.91	ug/l #	50
38) Carbon Tetrachloride	7.67	117	100586	4.68	ug/l #	17
43) Isopropyl Acetate	8.22	43	75219	3.49	ug/l #	65
47) Bromodichloromethane	9.41	83	6001	0.28	ug/l #	93
51) 4-Methyl-2-Pentanone	9.99	43	2742	0.24	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.54	63	29	14.33	ug/l #	51
64) Tetrachloroethene	10.63	164	23180	1.52	ug/l	97
65) Chlorobenzene	11.44	112	7753	0.20	ug/l	92
71) Bromoform	12.41	173	4306	0.46	ug/l #	1
74) N-amyl acetate	12.04	43	2049	0.23	ug/l #	47
75) 1,1,2,2-Tetrachloroethane	12.51	83	75	Below	Cal #	26
76) 1,2,3-Trichloropropane	12.40	75	298968	42.53	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	298968	134.43	ug/l #	9
84) 1,2,4-Trimethylbenzene	13.05	105	8598	0.30	ug/l	97
90) Hexachloroethane	13.90	117	112	Below	Cal #	9
95) Naphthalene	15.14	128	93033	10.05	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062218\
Data File : VN049504.D
Acq On : 22 Jun 2018 20:41
Operator : MD\SY
Sample : J3646-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jun 22 22:17:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion:168 Resp: 1131040

Ion Ratio Lower Upper

168 100

99 51.5 45.8 68.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

500000

400000

300000



200000

100000

0

Time-->

7.60 7.70 7.80



#8

Diethyl Ether

Concen: 0.92 ug/l

RT: 3.41 min Scan# 567

Delta R.T. 0.00 min

Lab File: VN049504.D


Acq: 22 Jun 2018 20:41

Tgt Ion: 74 Resp: 7129

Ion Ratio Lower Upper

74 100

45 103.0 51.9 155.8



Abundance Ion 74.00 (73.70 to 74.70): VNO

Ion 45.00 (44.70 to 45.70): VNO

3.41

3000


2000


1000

0

Time-->

3.35 3.40 3.45





#13

Acrolein

Concen: 0.28 ug/l

RT: 3.61 min Scan# 627

Delta R.T. 0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 56 Resp: 257

Ion Ratio Lower Upper

56 100

55 0.0 57.1 85.7#



#14

Allyl chloride

Concen: 1.52 ug/l

RT: 4.55 min Scan# 920

Delta R.T. 0.23 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41




Tgt Ion: 41 Resp: 31074


Ion Ratio Lower Upper

41 100

39 0.3 62.5 93.7#

76 0.0 27.1 40.7#









#16
Acetone
Concen: 19.32 ug/l
RT: 3.82 min Scan# 695
Delta R.T. 0.01 min
Lab File: VN049504.D
Acq: 22 Jun 2018 20:41




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 70360
Ion Ratio Lower Upper
43 100
58 34.3 22.4 33.6#



#18
Methyl Acetate
Concen: 7.51 ug/l
RT: 4.33 min Scan# 853
Delta R.T. 0.01 min
Lab File: VN049504.D
Acq: 22 Jun 2018 20:41

Tgt Ion: 43 Resp: 83352
Ion Ratio Lower Upper
43 100
74 23.3 17.4 26.0





#20

Methvlene Chloride

Concen: 57.87 ug/l

RT: 4.55 min Scan# 921

Delta R.T. 0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 84 Resp: 880368


Ion Ratio Lower Upper

84 100

49 119.8 106.9 160.3

51 35.5 32.7 49.1

86 65.7 49.8 74.8






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#25

2-Butanone

Concen: 0.24 ug/l

RT: 6.83 min Scan# 1631

Delta R.T. -0.01 min

Lab File: VN049504.D


Acq: 22 Jun 2018 20:41

Tgt Ion: 43 Resp: 1175

Ion Ratio Lower Upper



43 100


72 64.0 20.2 30.2#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 72.00 (71.70 to 72.70): VN0









#30
Chloroform
Concen: 0.53 ug/l
RT: 7.37 min Scan# 1799
Delta R.T. 0.00 min
Lab File: VN049504.D
Acq: 22 Jun 2018 20:41




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 83 Resp: 13766
Ion Ratio Lower Upper
83 100
85 64.2 51.7 77.5



#31
Cyclohexane
Concen: 0.79 ug/l
RT: 7.66 min Scan# 1889
Delta R.T. 0.01 min
Lab File: VN049504.D
Acq: 22 Jun 2018 20:41

Tgt Ion: 56 Resp: 21331
Ion Ratio Lower Upper
56 100
69 195.2 25.0 37.4#
84 143.4 64.6 96.8#





#33

1,2-Dichloroethane-d4

Concen: 52.04 ug/l

RT: 8.03 min Scan# 2003

Delta R.T. 0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 65 Resp: 780635

Ion Ratio Lower Upper

65 100

67 53.2 0.0 103.8



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0

8.03



300000

200000

100000

0

Time--> 7.90 8.00 8.10 8.20



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 8.59 min Scan# 2177

Delta R.T. 0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41


Tgt Ion: 114 Resp: 1810808

Ion Ratio Lower Upper

114 100

63 19.8 0.0 44.4

88 15.3 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0

8.59

1000000

800000


600000


400000

200000

0

Time--> 8.50 8.60 8.70 8.80









#35
 Dibromofluoromethane
 Concen: 47.03 ug/l
 RT: 7.59 min Scan# 1867
 Delta R.T. 0.00 min
 Lab File: VN049504.D
 Acq: 22 Jun 2018 20:41




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
113	100		
111	103.1	82.4	123.6
192	21.8	15.0	22.6



#36
 1,1-Dichloropropene
 Concen: 3.91 ug/l
 RT: 7.67 min Scan# 1891
 Delta R.T. -0.12 min
 Lab File: VN049504.D
 Acq: 22 Jun 2018 20:41

Tgt Ion	Ratio	Lower	Upper
75	100		
110	8.8	17.0	50.9#
77	0.0	24.4	36.6#





#38

Carbon Tetrachloride

Concen: 4.68 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. -0.11 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :
MSVOA_N
ClientSampled :


Tot Ion:117 Resp: 100586

Ion Ratio Lower Upper




117 100

119 5.7 75.6 113.4#

121 0.0 24.7 37.1#



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 3.49 ug/l

RT: 8.22 min Scan# 2063

Delta R.T. 0.06 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41


Tgt Ion: 43 Resp: 75219

Ion Ratio Lower Upper



43 100


61 1.2 14.2 21.4#

87 0.1 9.4 14.0#



Abundance Ion 43.00 (42.70 to 43.70): VNO
Ion 61.00 (60.70 to 61.70): VNO
Ion 87.00 (86.70 to 87.70): VNO





#47

Bromodichloromethane

Concen: 0.28 ug/l

RT: 9.41 min Scan# 2432

Delta R.T. 0.01 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 83 Resp: 6001

Ion Ratio Lower Upper

83 100

85 56.5 49.4 74.0

127 5.2 6.3 9.5#






Abundance Ion 82.90 (82.60 to 83.60): VN049504.D (-2337) (-)

Ion 84.90 (84.60 to 85.60): VN049504.D (-2337) (-)

Ion 126.80 (126.50 to 127.50): VN049504.D (-2337) (-)

Time-->



#50

Toluene-d8

Concen: 44.30 ug/l

RT: 10.09 min Scan# 2645

Delta R.T. 0.00 min

Lab File: VN049504.D


Acq: 22 Jun 2018 20:41

Tgt Ion: 98 Resp: 2470967

Ion Ratio Lower Upper

98 100



100 63.5 49.8 74.8




Abundance Ion 98.00 (97.70 to 98.70): VN049504.D (-2551) (-)

Ion 100.00 (99.70 to 100.70): VN049504.D (-2551) (-)

Time-->





#51

4-Methyl-2-Pentanone

Concen: 0.24 ug/l

RT: 9.99 min Scan# 2612

Delta R.T. 0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 2742

Ion Ratio Lower Upper

43 100

58 36.5 29.0 43.6



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0

2000

1500



1000

500

0

9.95 10.00 10.05

Time-->



#58

2-Chloroethyl Vinyl ether

Concen: 14.33 ug/l

RT: 9.54 min Scan# 2472

Delta R.T. -0.16 min

Lab File: VN049504.D


Acq: 22 Jun 2018 20:41

Tgt Ion: 63 Resp: 29

Ion Ratio Lower Upper

63 100

106 0.0 19.4 29.2#



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 105.90 (105.60 to 106.60): V

150


100


50

0

9.53 9.54 9.54

Time-->





#62

4-Bromofluorobenzene

Concen: 33.43 ug/l

RT: 12.41 min Scan# 3364

Delta R.T. 0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 95 Resp: 616537

Ion Ratio Lower Upper

95 100

174 89.8 0.0 165.6

176 86.7 0.0 157.4





Abundance Ion 94.90 (94.60 to 95.60): VN049263.D (-3347) (-)

Ion 173.80 (173.50 to 174.50): VN049263.D (-3347) (-)

Ion 175.80 (175.50 to 176.50): VN049263.D (-3347) (-)

12.41

Time--> 12.30 12.40 12.50



#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 11.41 min Scan# 3054

Delta R.T. 0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41


Tgt Ion: 117 Resp: 1489695

Ion Ratio Lower Upper

117 100

82 54.7 45.9 68.9

119 32.2 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): VN049263.D (-3038) (-)


Ion 82.00 (81.70 to 82.70): VN049263.D (-3038) (-)

Ion 118.90 (118.60 to 119.60): VN049263.D (-3038) (-)

11.41

Time--> 11.30 11.40 11.50





#64

Tetrachloroethene

Concen: 1.52 ug/l

RT: 10.63 min Scan# 2812

Delta R.T. -0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :

MSVOA_N

ClientSampleId :

Tot Ion:164 Resp: 23180





Ion Ratio Lower Upper

164 100

166 125.9 101.4 152.2

129 98.7 74.1 111.1

131 95.6 73.7 110.5



#65

Chlorobenzene

Concen: 0.20 ug/l

RT: 11.44 min Scan# 3062

Delta R.T. 0.00 min

Lab File: VN049504.D




Acq: 22 Jun 2018 20:41


Tgt Ion:112 Resp: 7753

Ion Ratio Lower Upper

112 100

114 28.6 26.2 39.4









#71
Bromoform
Concen: 0.46 ug/l
RT: 12.41 min Scan# 3364
Delta R.T. 0.28 min
Lab File: VN049504.D
Acq: 22 Jun 2018 20:41

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:173 Resp: 4306
Ion Ratio Lower Upper
173 100
175 738.2 24.5 73.5#
254 0.0 0.1 0.1#




Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.35 min Scan# 3657
Delta R.T. 0.00 min
Lab File: VN049504.D
Acq: 22 Jun 2018 20:41

Tgt Ion:152 Resp: 411944
Ion Ratio Lower Upper
152 100
115 56.6 28.7 86.1
150 157.7 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 0.23 ug/l

RT: 12.04 min Scan# 3251

Delta R.T. -0.03 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :

MSVOA_N

ClientSampleId :

Tot Ion: 43 Resp: 2049


Ion Ratio Lower Upper

43 100

70 0.0 30.3 45.5#

55 0.0 19.4 29.2#

61 3.0 17.8 26.6#






Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 70.00 (69.70 to 70.70): VNO

Ion 55.00 (54.70 to 55.70): VNO

Ion 61.00 (60.70 to 61.70): VNO



#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.51 min Scan# 3396

Delta R.T. 0.00 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41


Tgt Ion: 83 Resp: 75

Ion Ratio Lower Upper

83 100

131 0.0 5.0 15.0#



85 0.0 32.0 96.0#




Abundance Ion 82.90 (82.60 to 83.60): VNO

Ion 130.80 (130.50 to 131.50): VNO

Ion 84.90 (84.60 to 85.60): VNO





#76

1,2,3-Trichloropropane

Concen: 42.53 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion: 75 Resp: 298968

Ion Ratio Lower Upper

75 100

77 1.1 0.0 0.0#



#81

trans-1,4-Dichloro-2-butene

Concen: 134.43 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN049504.D

Acq: 22 Jun 2018 20:41




Tgt Ion: 75 Resp: 298968


Ion Ratio Lower Upper

75 100

53 0.2 82.7 124.1#

89 0.1 36.5 54.7#









#84
 1,2,4-Trimethylbenzene
 Concen: 0.30 ug/l
 RT: 13.05 min Scan# 3563
 Delta R.T. 0.00 min
 Lab File: VN049504.D
 Acq: 22 Jun 2018 20:41




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion:105 Resp: 8598
 Ion Ratio Lower Upper
 105 100
 120 46.3 22.2 66.6



#90
 Hexachloroethane
 Concen: Below Cal
 RT: 13.90 min Scan# 3827
 Delta R.T. 0.02 min
 Lab File: VN049504.D
 Acq: 22 Jun 2018 20:41

Tgt Ion:117 Resp: 112
 Ion Ratio Lower Upper
 117 100
 201 0.0 40.1 120.2#






#95
Naphthalene
Concen: 10.05 ug/l
RT: 15.14 min Scan# 4213
Delta R.T. 0.00 min
Lab File: VN049504.D
Acq: 22 Jun 2018 20:41

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
128	100		
127	13.3	10.6	16.0
129	12.5	8.5	12.7



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

