

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062218\
 Data File : VN049490.D
 Acq On : 22 Jun 2018 14:40
 Operator : MD\SY
 Sample : J3570-10RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20180373-AMENDED-COMPOSITE-

Quant Time: Jun 22 22:41:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1128670	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1693483	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1419731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	403475	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	741098	49.51	ug/l	0.00
Spiked Amount			Recovery	=	99.02%	
35) Dibromofluoromethane	7.59	113	505051	36.86	ug/l	0.00
Spiked Amount			Recovery	=	73.72%	
50) Toluene-d8	10.09	98	2381669	45.65	ug/l	0.00
Spiked Amount			Recovery	=	91.30%	
62) 4-Bromofluorobenzene	12.41	95	593346	34.40	ug/l	0.00
Spiked Amount			Recovery	=	68.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	127346	35.04	ug/l	93
20) Methylene Chloride	4.55	84	30492	2.01	ug/l	93
25) 2-Butanone	6.85	43	24103	4.94	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062218\
Data File : VN049490.D
Acq On : 22 Jun 2018 14:40
Operator : MD\SY
Sample : J3570-10RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20180373-AMENDED-COMPOSITE-

Quant Time: Jun 22 22:41:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

