

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062218\
 Data File : VN049499.D
 Acq On : 22 Jun 2018 18:32
 Operator : MD\SY
 Sample : PB116445ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116445ZHE#12

Quant Time: Jun 22 23:15:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1177965	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1845547	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1543241	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	384032	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	802995	51.40	ug/l	0.00
Spiked Amount			Recovery	=	102.80%	
35) Dibromofluoromethane	7.59	113	693819	46.47	ug/l	0.00
Spiked Amount			Recovery	=	92.94%	
50) Toluene-d8	10.09	98	2564992	45.12	ug/l	0.00
Spiked Amount			Recovery	=	90.24%	
62) 4-Bromofluorobenzene	12.41	95	601599	32.00	ug/l	0.00
Spiked Amount			Recovery	=	64.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	35077	9.25	ug/l #	85
18) Methyl Acetate	4.34	43	77973	6.74	ug/l	97
20) Methylene Chloride	4.55	84	50816	3.21	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062218\
Data File : VN049499.D
Acq On : 22 Jun 2018 18:32
Operator : MD\SY
Sample : PB116445ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB116445ZHE#12

Quant Time: Jun 22 23:15:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

