

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN061992.D
 Acq On : 22 Jun 2020 12:07
 Operator : JC/MD
 Sample : VN0622WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0622WBL01

Quant Time: Jun 22 15:05:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	160127	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	291434	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	281604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	104283	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	108711	50.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.24%	
35) Dibromofluoromethane	7.55	113	93184	49.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.96%	
50) Toluene-d8	10.06	98	367744	50.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.34%	
62) 4-Bromofluorobenzene	12.38	95	119311	47.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN061992.D
Acq On : 22 Jun 2020 12:07
Operator : JC/MD
Sample : VN0622WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0622WBL01

Quant Time: Jun 22 15:05:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

