

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN061998.D
 Acq On : 22 Jun 2020 14:41
 Operator : JC/MD
 Sample : L3043-02DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D1-20200615DL

Quant Time: Jun 23 05:41:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	141512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	257183	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	250907	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	106018	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	96501	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
35) Dibromofluoromethane	7.54	113	82294	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
50) Toluene-d8	10.06	98	326984	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
62) 4-Bromofluorobenzene	12.38	95	113883	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.71	101	6171	4.300	ug/l #	84
12) 1,1-Dichloroethene	3.69	96	2362	1.643	ug/l	87
44) Trichloroethene	8.80	130	97694	52.240	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN061998.D
Acq On : 22 Jun 2020 14:41
Operator : JC/MD
Sample : L3043-02DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D1-20200615DL

Quant Time: Jun 23 05:41:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

