

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN062000.D
 Acq On : 22 Jun 2020 15:30
 Operator : JC/MD
 Sample : L3043-23DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REDUP06-20200616DL

Quant Time: Jun 23 05:46:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	143475	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	264019	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	255013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	93427	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	101468	52.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.46%	
35) Dibromofluoromethane	7.55	113	86311	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
50) Toluene-d8	10.06	98	331234	49.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.76%	
62) 4-Bromofluorobenzene	12.38	95	106280	46.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.32%	
Target Compounds						
44) Trichloroethene	8.80	130	121889	63.490	ug/l	99
64) Tetrachloroethene	10.60	164	1083	0.666	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN062000.D
Acq On : 22 Jun 2020 15:30
Operator : JC/MD
Sample : L3043-23DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REDUP06-20200616DL

Quant Time: Jun 23 05:46:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

