

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN062004.D
 Acq On : 22 Jun 2020 17:06
 Operator : JC/MD
 Sample : L3033-05DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Z1-DUP-0571-200615DL

Quant Time: Jun 23 05:58:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	140191	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	256331	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	254396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	104306	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	96259	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
35) Dibromofluoromethane	7.55	113	80777	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	10.06	98	323281	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
62) 4-Bromofluorobenzene	12.38	95	112388	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
Target Compounds						
4) Vinyl Chloride	2.16	62	19522	8.415	ug/l	94
27) cis-1,2-Dichloroethene	6.79	96	3265	1.730	ug/l	96
40) Benzene	8.00	78	12300	1.737	ug/l	98
65) Chlorobenzene	11.41	112	256476	47.779	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN062004.D
Acq On : 22 Jun 2020 17:06
Operator : JC/MD
Sample : L3033-05DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Z1-DUP-0571-200615DL

Quant Time: Jun 23 05:58:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

