

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN062005.D
 Acq On : 22 Jun 2020 17:30
 Operator : JC/MD
 Sample : L3033-01DL 4X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS043MW100-200615DL

Quant Time: Jun 23 06:00:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	141767	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	263244	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	258224	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	106348	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	98790	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
35) Dibromofluoromethane	7.54	113	83781	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
50) Toluene-d8	10.06	98	330213	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
62) 4-Bromofluorobenzene	12.38	95	116104	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
Target Compounds						
4) Vinyl Chloride	2.16	62	24151	10.295	ug/l	98
27) cis-1,2-Dichloroethene	6.77	96	3637	1.905	ug/l	86
40) Benzene	8.00	78	16039	2.205	ug/l	97
65) Chlorobenzene	11.41	112	321828	59.065	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN062005.D
Acq On : 22 Jun 2020 17:30
Operator : JC/MD
Sample : L3033-01DL 4X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS043MW100-200615DL

Quant Time: Jun 23 06:00:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

