

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN062008.D
 Acq On : 22 Jun 2020 18:43
 Operator : JC/MD
 Sample : L3043-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D1-20200615

Quant Time: Jun 23 06:12:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	144640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	268224	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	260845	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	99136	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	101252	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
35) Dibromofluoromethane	7.55	113	88175	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
50) Toluene-d8	10.06	98	341862	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
62) 4-Bromofluorobenzene	12.38	95	111028	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.71	101	4062	2.769	ug/l	96
16) Acetone	3.79	43	2704	4.719	ug/l	90
27) cis-1,2-Dichloroethene	6.78	96	2956	1.518	ug/l	96
44) Trichloroethene	8.80	130	754059	386.620	ug/l	99
64) Tetrachloroethene	10.60	164	3325	2.000	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN062008.D
Acq On : 22 Jun 2020 18:43
Operator : JC/MD
Sample : L3043-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D1-20200615

Quant Time: Jun 23 06:12:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

