

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN062026.D
 Acq On : 23 Jun 2020 3:10
 Operator : JC/MD
 Sample : L3033-07 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LF001MW003-200616

Quant Time: Jun 23 07:08:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	160056	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	291192	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	282453	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	116221	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	106876	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
35) Dibromofluoromethane	7.55	113	91743	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
50) Toluene-d8	10.06	98	361497	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
62) 4-Bromofluorobenzene	12.38	95	128262	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
Target Compounds						
17) Carbon Disulfide	4.01	76	5013	1.054	ug/l	98
68) m/p-Xylenes	11.60	106	2147	0.557	ug/l	99
73) Isopropylbenzene	12.22	105	2575	0.312	ug/l	91
78) n-propylbenzene	12.57	91	5190	0.551	ug/l	99
80) 1,3,5-Trimethylbenzene	12.71	105	4058	0.594	ug/l	93
84) 1,2,4-Trimethylbenzene	13.01	105	14403	2.135	ug/l	99
95) Naphthalene	15.10	128	7198	4.641	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN062026.D
Acq On : 23 Jun 2020 3:10
Operator : JC/MD
Sample : L3033-07 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LF001MW003-200616

Quant Time: Jun 23 07:08:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

