

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN062033.D
 Acq On : 23 Jun 2020 5:58
 Operator : JC/MD
 Sample : PB129655ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129655ZHE#03

Quant Time: Jun 23 07:45:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	145196	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	282144	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	274688	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	104760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	102504	52.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.28%	
35) Dibromofluoromethane	7.55	113	87093	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
50) Toluene-d8	10.06	98	353457	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
62) 4-Bromofluorobenzene	12.38	95	117570	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
Target Compounds						
16) Acetone	3.78	43	8231	14.310	ug/l	96
18) Methyl Acetate	4.29	43	3490	2.107	ug/l #	65
20) Methylene Chloride	4.51	84	3144	1.678	ug/l	88
43) Isopropyl Acetate	8.13	43	39247	11.589	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN062033.D
Acq On : 23 Jun 2020 5:58
Operator : JC/MD
Sample : PB129655ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129655ZHE#03

Quant Time: Jun 23 07:45:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

