

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN062035.D
 Acq On : 23 Jun 2020 6:46
 Operator : JC/MD
 Sample : L3070-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 INF-SLUDGE

Quant Time: Jun 23 07:50:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	141621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	267217	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	261957	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	97267	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	97864	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
35) Dibromofluoromethane	7.55	113	84411	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
50) Toluene-d8	10.06	98	338468	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	12.38	95	110493	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
Target Compounds						
16) Acetone	3.78	43	19075	34.000	ug/l	99
18) Methyl Acetate	4.28	43	3548	2.197	ug/l #	84
20) Methylene Chloride	4.50	84	5636	3.083	ug/l	88
25) 2-Butanone	6.79	43	4503	5.501	ug/l	89
43) Isopropyl Acetate	8.13	43	25449	7.934	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN062035.D
Acq On : 23 Jun 2020 6:46
Operator : JC/MD
Sample : L3070-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
INF-SLUDGE

Quant Time: Jun 23 07:50:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

