

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062220\
 Data File : VN062036.D
 Acq On : 23 Jun 2020 7:11
 Operator : JC/MD
 Sample : L3081-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JFK-WC-S-SB-01

Quant Time: Jun 23 08:34:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	146253	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	273375	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	272233	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	101782	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	102576	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
35) Dibromofluoromethane	7.55	113	87993	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
50) Toluene-d8	10.06	98	352587	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
62) 4-Bromofluorobenzene	12.38	95	115468	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
Target Compounds						
16) Acetone	3.78	43	10880	18.779	ug/l	100
18) Methyl Acetate	4.28	43	2998	1.797	ug/l #	66
20) Methylene Chloride	4.50	84	6012	3.185	ug/l	99
43) Isopropyl Acetate	8.13	43	21475	6.544	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062220\
Data File : VN062036.D
Acq On : 23 Jun 2020 7:11
Operator : JC/MD
Sample : L3081-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JFK-WC-S-SB-01

Quant Time: Jun 23 08:34:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

