

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062221\
 Data File : VN067435.D
 Acq On : 22 Jun 2021 14:12
 Operator : JC/MD
 Sample : M2777-01ME
 Misc : 4.91g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 X-1ME

Quant Time: Jun 22 14:34:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	277251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	578533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	557454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	235586	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	268404	49.074	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.140%
35) Dibromofluoromethane	8.016	113	173013	44.965	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.940%
50) Toluene-d8	10.443	98	798419	54.139	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.280%
62) 4-Bromofluorobenzene	12.736	95	310461	51.011	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.020%
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.456	83	62244	9.068	ug/l	94
69) o-Xylene	12.283	106	77031	9.235	ug/l	98
80) 1,3,5-Trimethylbenzene	13.063	105	277718	14.584	ug/l	100
84) 1,2,4-Trimethylbenzene	13.372	105	468921	24.925	ug/l	100
85) sec-Butylbenzene	13.503	105	344700	16.056	ug/l	99
86) p-Isopropyltoluene	13.619	119	190146	10.974	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062221\
Data File : VN067435.D
Acq On : 22 Jun 2021 14:12
Operator : JC/MD
Sample : M2777-01ME
Misc : 4.91g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
X-1ME

Quant Time: Jun 22 14:34:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

