

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062221\
 Data File : VN067445.D
 Acq On : 22 Jun 2021 18:27
 Operator : JC/MD
 Sample : M2764-02RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PHC-250EM-001RE

Quant Time: Jun 23 04:06:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	284710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	594827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	532635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	181551	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	259538	46.210	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.420%
35) Dibromofluoromethane	8.024	113	180125	45.531	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.060%
50) Toluene-d8	10.446	98	781436	51.536	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.080%
62) 4-Bromofluorobenzene	12.734	95	264056	42.198	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.400%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	64836	30.081	ug/l	93
20) Methylene Chloride	5.101	84	25640	5.717	ug/l	94
25) 2-Butanone	7.351	43	20407	6.626	ug/l	93
27) cis-1,2-Dichloroethene	7.327	96	21536	4.837	ug/l	80
43) Isopropyl Acetate	8.568	43	44985	3.547	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062221\
Data File : VN067445.D
Acq On : 22 Jun 2021 18:27
Operator : JC/MD
Sample : M2764-02RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PHC-250EM-001RE

Quant Time: Jun 23 04:06:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

