

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062222\
 Data File : VN073022.D
 Acq On : 22 Jun 2022 13:48
 Operator : JC\MD
 Sample : VN0622WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0622WBL01

Quant Time: Jun 23 07:31:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	313014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	508036	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	459094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	223442	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	187037	44.197	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.400%
35) Dibromofluoromethane	7.951	113	161065	51.293	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.580%
50) Toluene-d8	10.375	98	523794	44.828	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.660%
62) 4-Bromofluorobenzene	12.669	95	225408	53.649	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062222\
Data File : VN073022.D
Acq On : 22 Jun 2022 13:48
Operator : JC\MD
Sample : VN0622WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0622WBL01

Quant Time: Jun 23 07:31:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

