

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062222\
 Data File : VN073026.D
 Acq On : 22 Jun 2022 15:41
 Operator : JC\MD
 Sample : N3381-17DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PZ-9DL

Quant Time: Jun 23 07:32:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	288648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	465072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	406803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	197904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	169964	43.553	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.100%		
35) Dibromofluoromethane	7.951	113	146903	51.104	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.200%		
50) Toluene-d8	10.374	98	478238	44.711	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	89.420%		
62) 4-Bromofluorobenzene	12.668	95	201224	52.317	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.640%		
Target Compounds						
					Qvalue	
16) Acetone	4.245	43	5318	2.479	ug/l #	67
31) Cyclohexane	8.022	56	11470	2.173	ug/l #	58
40) Benzene	8.392	78	68760	5.625	ug/l	99
52) Toluene	10.439	92	14409	1.920	ug/l	99
67) Ethyl Benzene	11.780	91	265494	18.984	ug/l	100
68) m/p-Xylenes	11.886	106	326077	59.445	ug/l	97
69) o-Xylene	12.216	106	2273	0.415	ug/l	92
73) Isopropylbenzene	12.515	105	12030	0.786	ug/l	97
78) n-propylbenzene	12.857	91	15064	0.895	ug/l	98
80) 1,3,5-Trimethylbenzene	12.992	105	5834	0.464	ug/l	95
84) 1,2,4-Trimethylbenzene	13.304	105	8631	0.689	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062222\
Data File : VN073026.D
Acq On : 22 Jun 2022 15:41
Operator : JC\MD
Sample : N3381-17DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PZ-9DL

Quant Time: Jun 23 07:32:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

