

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062222\
 Data File : VN073034.D
 Acq On : 22 Jun 2022 18:55
 Operator : JC\MD
 Sample : N3381-11 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-24

Quant Time: Jun 23 07:33:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	298385	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	489831	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	428152	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	195055	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	179390	44.469	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.940%
35) Dibromofluoromethane	7.951	113	154027	50.874	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.740%
50) Toluene-d8	10.375	98	506584	44.967	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.940%
62) 4-Bromofluorobenzene	12.669	95	207121	51.128	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.260%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
40) Benzene	8.392	78	1882346	146.203	ug/l	98
52) Toluene	10.439	92	39104	4.947	ug/l	98
67) Ethyl Benzene	11.780	91	372730	25.322	ug/l	99
68) m/p-Xylenes	11.886	106	450464	78.027	ug/l	98
69) o-Xylene	12.216	106	49845	8.640	ug/l	99
73) Isopropylbenzene	12.516	105	5111	0.339	ug/l	98
78) n-propylbenzene	12.857	91	9168	0.552	ug/l	94
80) 1,3,5-Trimethylbenzene	12.992	105	11999	0.968	ug/l	96
84) 1,2,4-Trimethylbenzene	13.304	105	52931	4.287	ug/l	98
95) Naphthalene	15.433	128	33795	2.243	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062222\
Data File : VN073034.D
Acq On : 22 Jun 2022 18:55
Operator : JC\MD
Sample : N3381-11 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-24

Quant Time: Jun 23 07:33:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

