

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062222\
 Data File : VN073035.D
 Acq On : 22 Jun 2022 19:19
 Operator : JC\MD
 Sample : N3381-12 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-26

Quant Time: Jun 23 07:34:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	283855	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	461663	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	411446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	185542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	171645	44.727	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.460%
35) Dibromofluoromethane	7.951	113	146854	51.465	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.920%
50) Toluene-d8	10.375	98	476372	44.865	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.720%
62) 4-Bromofluorobenzene	12.669	95	197570	51.746	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.500%
Target Compounds						
					Qvalue	
40) Benzene	8.392	78	3622	0.298	ug/l	97
67) Ethyl Benzene	11.786	91	3137	0.222	ug/l #	82
68) m/p-Xylenes	11.886	106	3525	0.635	ug/l	84
84) 1,2,4-Trimethylbenzene	13.304	105	5051	0.430	ug/l	90
95) Naphthalene	15.433	128	7290	0.509	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062222\
Data File : VN073035.D
Acq On : 22 Jun 2022 19:19
Operator : JC\MD
Sample : N3381-12 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-26

Quant Time: Jun 23 07:34:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

